Ring Communications (UK) Ltd
Newmarket Business Centre
341 Exning Road
Newmarket, Suffolk, CB8 0AT
Tel: +44(0) 1638 660 629
E-Mail sales@ringuk.com
Web: www.ringuk.co.uk

Cath-Lab to Control Room Intercom System – Pulse



Installation
Desk/wall station should be mounted with easy access to controls.
System Control Box can be mounted in the ceiling, but may require access for maintenance.
Cable between Desk/wall Station and the System Control Box is CAT5 cable.
Cable for microphone should be microphone cable, 2 core plus Screen.
The speaker requires 2 core speaker cable.

Recommendations
The box for mounting the Cardioid Condenser Microphone can be fitted to the bottom of monitor frame, using double sided sticky foam.
Position the Cardioid Microphone to be pointing at the operators, try to avoid equipment noise.
Cable for microphone needs running down the cable route for monitor stack. (Normally inside cable stocking)
The microphone must point at the person for best pick up. Use the 'knuckle-joint' hinge for positioning and direction of the microphone head.
The unique knuckle-joint provides both horizontal rotation and vertical angle adjustment.

Operation
To communicate with the Catheter Lab, press the Star key "*", then press and hold the "M" key to talk to the patient, let go to return to listen mode.

Patient in the Cath-lab, cannot hear operator without the "M" key being pressed.

Close System
To cancel all communications (Switch OFF) press "C".